

USER MANUAL

DICC-35

CONDUCTIVITY/TDS CONTROLLER

SINGLE CHANNEL






Ver. 1.0









SPECIFICATIONS

POWER	: 220~240VAC , 50HZ
CONDUCTIVITY RANGE	: 0~10uS
CONDUCTIVITY RESOLUTION	: 0~0.01uS
CONDUCTIVITY ACCURACY	: \pm 0.1uS
CONDUCTIVITY DIGITAL OUTPUT	: 5A, 1CO, RELAY
CONDUCTIVITY ANALOG OUTPUT	: 4~20mA
CELL CONSTANT	: 0.1k
DISPLAY	: 3.5" Color LCD
BODY	: ABS
OUTER DIMENSIONS	: 96x96x110mm
CUTOUT DIMENSIONS	: 92x92mm

KEY FUNCTION

	TO SET TDS CONTROLLING VALUE
	TO INCREASE VALUE & IT WILL SCROLL 0 TO 9
	TO SHIFT DIGIT
	+  TO SAVE DATA &EXIT TO MAIN DISPLAY

CALIBRATION KEY FUNCTION

	+ 	HOLD FOR 5 SECOND
		TO INCREASE VALUE & IT WILL SCROLL - OR + & 0 TO 9
		TO SHIFT DIGIT
	+ 	TO SAVE DATA &EXIT TO MAIN DISPLAY

ANALOG OUTPUT

ANALOG OUT PUT : 4~20mA [FOUR WIRE]

SR NO: _____



DATA INSTRUMENTS



DATA INSTRUMENTS

Reg. Address :
16, DHARMADEVNAGAR,
Opp. SAMARATNAGAR,
ISANPUR, Ahmedabad - 382443.

FACTORY :
61, Sankheshwar Industrial Estate,
Opp. NEO Plast,
B/H. HDFC Bank, Phase - 1, GIDC,
Vatva, Ahmedabad - 382445.

Web : www.datainstrumentsindia.com
e-mail : utkarsh@datainstrumentsindia.com, contact@datainstrumentsindia.com
Cont. : (F) 09909035088, +91 93288 88879